



Certificate of Analysis

Pesticides Mixture

Product Number: US-118

Page: 1 of 1

Lot Number: CK-2686

Lot Issue Date: 15-Aug-2013

Expiration Date: 31-Mar-2014

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aramite (total)	000140-57-8	RM06550	2006 ± 10 µg/mL
chlorobenzilate	000510-15-6	NT00526	2000 ± 10 µg/mL
diallate (total)	002303-16-4	RM06660	2000 ± 10 µg/mL
isodrin	000465-73-6	RM03076	1995 ± 10 µg/mL
kepone	000143-50-0	RM00116	2006 ± 10 µg/mL
pronamide	023950-58-5	RM03337	2007 ± 10 µg/mL

Matrix: methylene chloride (dichloromethane)

Storage: Store at Room Temperature (15-30°C)

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2008
Registered
TUV USA, Inc.
Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager