



Certificate of Analysis

Pesticides Mixture

Product Number: US-118

Page: 1 of 1

Lot Number: CM-4895

Lot Issue Date: 16-Sep-2015

Expiration Date: 30-Apr-2016

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aramite (total)	000140-57-8	RM06550	2000 ± 10 µg/mL
chlorobenzilate	000510-15-6	RM07447	2002 ± 10 µg/mL
diallate (total)	002303-16-4	RM09218	2002 ± 10 µg/mL
isodrin	000465-73-6	RM03076	2009 ± 10 µg/mL
kepone	000143-50-0	RM03986	2006 ± 10 µg/mL
pronamide	023950-58-5	RM07251	2004 ± 10 µg/mL

Matrix: methylene chloride (dichloromethane)

Storage: Store at Room Temperature (15° to 30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 9001
Registered
TUV USA, Inc.


William J. Leary
Quality Assurance Manager