

Certificate of Analysis

ISO Guide 34

Pesticides Standard


Product Number: US-118**Page:** 1 of 1**Lot Number:** CT-0947**Lot Issue Date:** 28-Feb-2019**Expiration Date:** 30-Sep-2019

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
aramite (total)	000140-57-8	RM09669	2005 ± 10 µg/mL
chlorobenzilate	000510-15-6	RM14480	2000 ± 10 µg/mL
diallate (total)	002303-16-4	RM14453	2002 ± 10 µg/mL
isodrin	000465-73-6	RM03076	2001 ± 10 µg/mL
kepone	000143-50-0	RM03986	2002 ± 10 µg/mL
pronamide	023950-58-5	RM14132	2000 ± 10 µg/mL

Matrix: methylene chloride (dichloromethane)**Storage:** Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS RepresentativeISO Guide 34 Cert No.
AR-1936Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management SystemISO17025 Cert No.
AT-1937