



## CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLY-SA™ SIALIDASE A™  
PRODUCT CODE: WS0376  
LOT NUMBER: W180055  
STORAGE: 2°C - 8°C  
EXPIRATION: January 2020; the components of this pack are usable at least through this month.

QC RELEASE:

The Sialidase A was tested with WS0379 Gly-SA SAQ Reagent Pack and conforms to specifications.<sup>1</sup>

---

Authorized Signature

<sup>1</sup> WS0386 Bovine Fetuin (0.4 mg) was dissolved in 1.6 ml of deionized water to a concentration of 0.25 mg/ml. A 10 µl (2.5 µg) aliquot of the diluted fetuin digested with 0.05 U (10 µl) of WS0376 Sialidase A™ for 30 minutes at 37°C yielded 0.20 – 0.29 nmols of sialic acid when processed as per the GS96-SAQ kit protocol using fluorescence detection.