



Agilent Technologies



Certificate of Analysis

Test Std, DB-1/DB-5 (Microbore/HiTemp)

**Agilent Part
Number:** 200-0010

**Sample Lot
Number:** CH-3304

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The certified value and uncertainty value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	True Value
n-decane	000124-18-5	9401284	0.251 ± 0.001 mg/mL
n-tridecane	000629-50-5	2290690	0.251 ± 0.001 mg/mL
1-decanol	000112-30-1	032057	0.251 ± 0.001 mg/mL
1-octanol	000111-87-5	RM02416	0.251 ± 0.001 mg/mL
2,6-dimethylphenol	000576-26-1	RM02150-01	0.251 ± 0.001 mg/mL
2,6-dimethylaniline	000087-62-7	RM00428-01	0.251 ± 0.001 mg/mL
naphthalene	000091-20-3	RM02406-02	0.251 ± 0.001 mg/mL
methyl decanoate	000110-42-9	NT00187	0.251 ± 0.001 mg/mL

Solvent: hexane

ULTRA Scientific uses calibrated Class A glassware and clean bottles in the manufacture of ULTRA standards. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: 17 October 2011
Date of expiration: 30 November 2013


William J. Leary
Quality Assurance Manager