



Certificate of Analysis

Test Std, DB-FFAP/OV-351 (Capillary)

**Agilent Part
Number:** 200-0032

**Sample Lot
Number:** CP-4893

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.


Analyte	CAS#	Analyte Lot	Certified Value
n-tetradecane	000629-59-4	RM09997	0.251 ± 0.001 mg/mL
n-tridecane	000629-50-5	RM04185	0.251 ± 0.001 mg/mL
1-octanol	000111-87-5	RM02416	0.251 ± 0.001 mg/mL
hexanoic acid	000142-62-1	NT01887	0.251 ± 0.001 mg/mL
isovaleric acid	000503-74-2	RM00254	0.250 ± 0.001 mg/mL
2,6-dimethylphenol	000576-26-1	RM02150	0.250 ± 0.001 mg/mL
naphthalene	000091-20-3	RM02406	0.251 ± 0.001 mg/mL
methyl decanoate	000110-42-9	NT00187	0.250 ± 0.001 mg/mL
methyl dodecanoate	000111-82-0	NT01095	0.251 ± 0.001 mg/mL
methyl undecanoate	001731-86-8	NT00201	0.251 ± 0.001 mg/mL


Solvent: hexane

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: 28 September 2016

Date of expiration: 31 October 2018


John Russo
President


Monica Bourgeois
Director of QA/RA