Agilent Technologies

Certificate of Analysis

Test Std, DB-624 (Cap/Mega)

Agilent Part Number: **200-0113**

Sample Lot Number: **CS-2896**

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Certified Value
n-heptane	000142-82-5	RM11823	1.25 ± 0.01 mg/mL
n-nonane	000111-84-2	RM11940	2.00 ± 0.01 mg/mL
n-octane	000111-65-9	NT01970	1.40 ± 0.01 mg/mL
methanol	000067-56-1	RM11016	1.27 ± 0.01 mg/mL
1-propanol	000071-23-8	RM05295	1.25 ± 0.01 mg/mL
pyridine	000110-86-1	RM06905	1.41 ± 0.01 mg/mL
2-butanone	000078-93-3	RM06696	1.25 ± 0.01 mg/mL
acetone	000067-64-1	RM11831	1.26 ± 0.01 mg/mL
1,2-dichloropropane	000078-87-5	RM12821	2.01 ± 0.01 mg/mL
tetrachloroethylene	000127-18-4	RM06491	3.24 ± 0.02 mg/mL
chlorobenzene	000108-90-7	RM01874	2.00 ± 0.01 mg/mL

Solvent: methylene chloride (dichloromethane)

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: **01 June 2018**
Date of expiration: **30 June 2020**

John Russo
President

Monica Bourgeois
Director of QA/RA