Agilent Technologies SIGMA-ALDRICH®

Certificate of Analysis

Description : Amino Acid Standard (0.025 nmol/µl) 10/pk

Part.-No. : 5061-3333 Production date : 20-Mar 2018
Lot-No. : BCBW7955 Expiration date : 20-Mar 2020

No.	Component	Molecular Weight	Final Concentration (mg/ml)
01	**L-Alanine**	89.10	0.00223
02	**L-Arginine**	174.20	0.00436
03	**L-Aspartic Acid**	133.11	0.00333
04	**L-Cystine**	240.30	0.00601
05	**L-Glutamic Acid**	147.13	0.00368
06	**Glycine**	75.07	0.00188
07	**L-Histidine Hydrochloride Monohydrate**	209.63	0.00524
08	**L-Isoleucine**	131.18	0.00328
09	**L-Leucine**	131.18	0.00328
10	**L-Lysine Hydrochloride**	182.65	0.00457
11	**L-Methionine**	149.21	0.00373
12	**L-Phenylalanine**	165.19	0.00413
13	**L-Proline**	115.13	0.00288
14	**L-Serine**	105.09	0.00263
15	**L-Threonine**	119.12	0.00298
16	**L-Tyrosine**	181.19	0.00453
17	**L-Valine**	117.15	0.00293

Purity Determination:
Amino acids by Titration, Loss on drying, Residue on ignition, IR-Spectroscopy, MicroSelect-Test, Trace Metal and Trace Anion Determination
Hydrochloric acid by Titration, MicroSelect-Test, Trace Metal and Trace Anion Determination

Raw materials
All raw materials used to prepare this amino acid standard are of the highest available purity (> 99%) and are routinely analyzed according to the above-mentioned purity-determinations.

Manufacturing
We employ precise measuring techniques in manufacturing this amino acid standard. Mass is determined with electronic balances capable of weighing to 0.0001 g and calibrated by the Swiss Office of Weights and Measures. Volume ist determined in dedicated high-purity borosilicate volumetric flasks capable of measuring a 2000 ml-volume with an accuracy of 0.3%.

Packaging and Storage
The final solution is handled under argon, filled into 1 ml amber ampoules under inert gas (argon) and sealed. 10 finished ampoules are packaged into a set and stored at 4 °C.

Stability
Every individual lot of the product is subjected to reanalysis and the experience allows to set the shelf life to two years, if the product is stored as received at 4 °C. The guaranteed stability is not applicable to ampoules stored after opening, even if resealed.

Analytical Quality Control
The scope of the analytical testing procedures covers identity, purity, homogeneity, accuracy, function test in amino acid analysis and stability of the finished product

Description	Lot Analysis	Specifications
Aspect	clear, colorless liquid	clear, colorless liquid
Density (20/4)	1.000	1.00 ± 0.01 g/ml
Index of Refraction (20/D)	1.334	1.334 ± 0.005
Amino Acid Analysis :	corresponds	corresponds
- Identity	corresponds	corresponds
- Purity	corresponds	corresponds
- Concentration Accuracy	corresponds	± 2.5% relative to primary standard

Buchs, 20.03.2018

Sigma-Aldrich Production GmbH
Industriestrasse 25
CH-9470 Buchs / Switzerland

A. Schneider

Dr. A. Schneider, Chemist Quality Assurance Sr
Quality Assurance, Buchs / Switzerland

Sigma-Aldrich warrants that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product for its particular use. See invoice for additional terms and conditions of sale.