



## Certificate of Analysis

Description : OPA-Reagent, 10 mg/ml, 6 ampoules  
Part.-No. : 5061-3335 Production date : 19 Dec 2022  
Lot-No. : BCCJ8541 Expiration date : 19-Dec 2023

### Raw materials

All raw materials used to prepare this OPA-Reagent are of the highest available purity and are routinely analyzed according to the specified purity-determinations.

### Manufacturing

We employ precise measuring techniques in manufacturing. Mass is determined with electronic balances capable of weighing to 0.0001 g and calibrated by the Swiss Office of Weights and Measures. Volume is determined in dedicated high-purity borosilicate volumetric flasks capable of measuring with an accuracy of 0.3%.

### Packaging and Storage

The final solution was handled under argon and automatically sealed in 1 ml amber ampoules. Finished ampoules are packaged and stored at 2 – 8 °C.

### Stability

Every individual lot of the product is subjected to reanalysis and the experience allows to set the shelf life to one year, if the product is stored as received at 2 – 8 °C. The guaranteed stability is not applicable to ampoules stored after opening, even if resealed.

### Analytical Quality Control

The scope of the analytical testing procedures covers identity, purity, homogeneity, accuracy, function test in amino acid analysis and stability of the finished product

Description	Lot Analysis	Specifications
Aspect	clear, yellowish liquid	clear, yellowish liquid
Density (20/4)	1.044	1.04 ± 0.01 g/ml
Index Of Refraction (20/D)	1.349	1.348 ± 0.005
Amino Acid Analysis :	corresponds	standard and blank correspond
- Identity	corresponds	corresponds
- Purity	corresponds	corresponds
- Concentration Accuracy	corresponds	± 5 %
Luminescence Test	corresponds	corresponds