

Certificate of Analysis

ICP-MS Stock Tuning Solution (100 mL)

Agilent Part Number: 5188-6564
Lot Number: 20-122VYY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 mg/L	10.2 mg/L	3110*	Ce(NO ₃) ₂ · 6H ₂ O	99.99+	Tl	7440-28-0	10.0 mg/L	10.0 mg/L	3158*	TlNO ₃	99.99+
Co	7440-48-4	10.0 mg/L	10.1 mg/L	3113*	Co	99.99+	Y	7440-65-5	10.0 mg/L	9.96 mg/L	3167a*	Y ₂ O ₃	99.99+
Li	7439-93-2	10.0 mg/L	9.98 mg/L	3129a*	Li ₂ CO ₃	99.99+							

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

 2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 19-66VY, 19-67VY, 19-245VY, 19-246VY.

Trace Metallic Impurities in the Actual Solution, in µg/L, via ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	<0.09	Cs	<0.03	Ho	<0.01	Nd	<0.03	Ru	<0.6	Te	<0.7
Al	0.4	Cu	<0.5	In	<0.04	Ni	<0.2	Sb	<0.3	Th	<0.05
As	<1	Dy	<0.02	Ir	<0.4	P	<50	Sc	<0.2	Ti	<0.8
Au	<0.6	Er	<0.02	K	<70	Pb	<0.5	Se	<2	Tm	<0.01
B	<0.8	Eu	<0.01	La	0.5	Pd	<2	Si	<100	U	<0.02
Ba	<0.3	Fa	5	Lu	<0.01	Pr	0.2	Sm	<0.02	V	<0.2
Be	<0.03	Ga	0.4	Mg	<0.5	Pt	<0.2	Sn	<0.1	W	<1
Bi	<0.04	Gd	<0.05	Mn	<0.5	Rb	<0.07	Sr	<0.05	Yb	<0.01
Ca	20	Ge	<0.07	Mo	<0.09	Re	<0.02	Ta	<0.07	Zn	<0.4
Cd	<0.03	Hf	<0.01	Na	<10	Rh	<0.3	Tb	<0.1	Zr	<0.4
Cr	<0.3	Hg	<0.1	Nb	<0.05						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A.

Date of release: August 31, 2013

Date of expiration: February 28, 2015

Brijender S. Tonk
 QC Coordinator
 CertiPrep, Inc.

 Visit Agilent on the world wide web: www.agilent.com/chem/supplies