



Certificate of Analysis

ICP-MS Stock Tuning Solution (100 mL)

Agilent Part Number: 5188-6564

Lot Number: 8-140MFY2

Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity	Analyte	CAS#	Labeled Conc.	Measured Conc.	SRM	Start Mat'l Formula	Start Mat'l Purity
Ce	7440-45-1	10.0 µg/mL	10.1 µg/mL	3110*	Ce ₂ (CO ₃) ₃ ·x(H ₂ O)	99.99+	Tl	7440-28-0	10.0 µg/mL	9.99 µg/mL	3158*	TlNO ₃	99.99+
Co	7440-48-4	10.0 µg/mL	10.0 µg/mL	3113*	Co	99.99+	Y	7440-65-5	10.0 µg/mL	9.98 µg/mL	3167a*	Y ₂ O ₃	99.99+
Li	7439-93-2	10.0 µg/mL	9.96 µg/mL	3129a*	Li ₂ CO ₃	99.99+							

* - indicates NIST SRM

† - indicates CRM (when NIST SRM is not available)

Purity grades:

Starting Materials: Shown above

Matrix:

2% HNO₃: HNO₃ (CAS No. 7697-37-2) high purity grade

Traceability:

This standard has been produced gravimetrically and volumetrically using ISO 9001 quality procedures. Agilent ICP / ICP-MS Spectrometer was used to determine the concentration of the main elements via NIST SRMs shown above, as well as the impurities. Other reference standards used: 56-137CR, 3-210SG, 7-246MF.

Trace Metallic Impurities in the Actual Solution, in µg/L, via Agilent ICP-MS Analysis, results are accurate to ±10%:

Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.	Element	Conc.
Ag	0.05	Cs	<0.02	Ho	<0.01	Nd	<0.01	Ru	0.2	Te	<0.2
Al	1	Cu	<0.05	In	0.03	Ni	<0.01	Sb	0.02	Th	0.01
As	<0.4	Dy	<0.02	Ir	<0.01	P	<100	Sc	<0.01	Ti	<0.01
Au	<0.06	Er	0.2	K	<30	Pb	<0.04	Se	<0.1	Tm	<0.01
B	0.5	Eu	<0.02	La	<0.01	Pd	0.08	Si	<100	U	<0.01
Ba	0.06	Fe	1	Lu	<0.01	Pr	0.2	Sm	<0.02	V	<0.06
Be	<0.08	Ga	0.4	Mg	<0.3	Pt	0.01	Sn	<0.05	W	<0.02
Bi	<0.02	Gd	0.2	Mn	<0.2	Rb	<0.04	Sr	<0.07	Yb	<0.01
Ca	0.5	Ge	<0.09	Mo	<0.04	Re	<0.02	Ta	<0.01	Zn	<0.6
Cd	<0.01	Hf	<0.01	Na	<2	Rh	<0.01	Tb	1	Zr	<0.01
Cr	<0.5	Hg	<0.07	Nb	<0.01						

Balances are calibrated regularly with weight sets traceable to NIST.

Agilent reference standards are guaranteed stable and accurate to ±0.5% of measured analyte concentration. This uncertainty is at 95% confidence interval, a coverage factor of 2. For these solutions we use the highest purity acids applicable, 18 megohm double deionized water and acid-leached, triple rinsed bottles. All glassware used is class A. This standard was manufactured following the guidelines set forth under ISO 17025 and ISO 17034 regulations.

Date of release: November 30, 2022

Date of expiration: May 31, 2024

Brijender S Tonk
 QC Coordinator