



Certificate of Analysis

Pesticide Analyzer Checkout Solution

**Agilent Part
Number:** 5190-0468

**Sample Lot
Number:** CM-1729

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Purity %	Certified Value
dichlorvos	000062-73-7	RM07626	98	10.0 ± 0.1 µg/mL
mevinphos	007786-34-7	RM07082	95	10.0 ± 0.1 µg/mL
ethalfluralin	055283-68-6	RM03918	98	10.0 ± 0.1 µg/mL
trifluralin	001582-09-8	RM05766	99	10.0 ± 0.1 µg/mL
atrazine	001912-24-9	RM03843	99	10.0 ± 0.1 µg/mL
lindane	000058-89-9	RM02694	99	9.99 ± 0.05 µg/mL
chlorpyrifos-methyl	005598-13-0	NT02119	99	10.0 ± 0.1 µg/mL
heptachlor	000076-44-8	RM03111	99	10.0 ± 0.1 µg/mL
malathion	000121-75-5	RM07001	98	10.0 ± 0.1 µg/mL
chlorpyrifos	002921-88-2	RM09586	99	10.0 ± 0.1 µg/mL
dieltrin	000060-57-1	RM03957	98	10.0 ± 0.1 µg/mL
4,4'-DDE	000072-55-9	RM02891	99	10.0 ± 0.1 µg/mL
hexazinone	051235-04-2	RM03124	99	10.0 ± 0.1 µg/mL
propargite	002312-35-8	RM03259	96	10.0 ± 0.1 µg/mL
leptophos	021609-90-5	RM03132	99	10.0 ± 0.1 µg/mL
mirex	002385-85-5	RM07019	99	10.0 ± 0.1 µg/mL
fenarimol	060168-88-9	RM02969	99	10.0 ± 0.1 µg/mL
coumaphos	000056-72-4	RM07758	99	10.0 ± 0.1 µg/mL
ethofenprox	080844-07-1	RM03732	99	10.0 ± 0.1 µg/mL
deltamethrin	052918-63-5	RM07731	99	10.0 ± 0.1 µg/mL

Solvent: acetone

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


Date of release: 23 April 2015

Date of expiration: 31 May 2016


William J. Lea
Quality Assurance Manager



Agilent Technologies




William J. Lea
Quality Assurance Manager