



Certificate of Analysis

HPLC Internal Standard, EN Method 15662

**Agilent Part
Number:** 5190-0500

**Sample Lot
Number:** CP-2306

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Certified Value
tris(1,3-dichloroisopropyl)phosphate	013674-87-8	RM01914	100.3 ± 0.5 µg/mL
nicarbazin	000330-95-0	RM02436	99.8 ± 0.5 µg/mL


Solvent: acetonitrile

The product quality was confirmed by analysis with GC-MS.

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: 12 May 2016

Date of expiration: 30 June 2018


Peter A. King, Ph.D.
VP, Technical Operations


Daniel J. Lamendola
Director of QA/RA