



Certificate of Analysis

LC/MS Pesticide Standard #6

Part Number: 5190-0551-6**Lot Number:** CS-3128**Date of release:****06 June 2018****Date of expiration:****31 July 2019**

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Purity %	Certified Value
carbaryl	000063-25-2	RM09161	99.9	100.0 ± 0.5 µg/mL
trichlorfon	000052-68-6	RM03381	98.0	100.4 ± 0.5 µg/mL
propoxur	000114-26-1	RM11363	99.9	100.1 ± 0.5 µg/mL
trietazine	001912-26-1	RM03941	99.6	100.2 ± 0.5 µg/mL
2,3,5-trimethacarb	002655-15-4	RM11353	97.3	100.2 ± 0.5 µg/mL
carbosulfan	055285-14-8	RM12456	98.5	100.3 ± 0.5 µg/mL
diethofencarb	087130-20-9	RM10618	99.5	100.0 ± 0.5 µg/mL
fenoxycarb	072490-01-8	RM12794	98.8	100.3 ± 0.5 µg/mL
furathiocarb	065907-30-4	RM13653	98.0	100.1 ± 0.5 µg/mL
pencycuron	066063-05-6	RM10738	98.4	100.0 ± 0.5 µg/mL
pyrimethanil	053112-28-0	RM09179	99.9	100.2 ± 0.5 µg/mL
triadimenol	055219-65-3	RM03453	98.0	100.5 ± 0.5 µg/mL
flonicamid	158062-67-0	RM12972	99.5	100.4 ± 0.5 µg/mL
flumioxazin	103361-09-7	RM13092	99.1	100.2 ± 0.5 µg/mL
spiromesifen	283594-90-1	RM09770	99.5	100.5 ± 0.5 µg/mL
spirotramat	203313-25-1	RM11902	99.5	100.0 ± 0.5 µg/mL
sulfentrazone	122836-35-5	RM13708	99.5	100.3 ± 0.5 µg/mL
zoxamide	156052-68-5	RM09337	98.0	100.4 ± 0.5 µg/mL
omethoate	001113-02-6	RM10350	96.8	100.5 ± 0.5 µg/mL
mesosulfuron-methyl	208465-21-8	RM03233	97.5	100.1 ± 0.5 µg/mL
phosmet	000732-11-6	RM03255	99.0	100.4 ± 0.5 µg/mL

Solvent: acetonitrile**Storage:** Store Frozen (-25° to -10°C).

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


John Russo
President
Monica Bourgeois
Director of QA/RA