



Agilent Technologies
www.agilent.com

Expiry: _____

Certificate of Analysis

Part Number: 6610030500
Lot Number: 11D258
Shelf Life: 18 Months

Calibration Mix 1
2% HNO₃ + 0.5% HF

Concentrations in ug/mL \pm 0.5%

Sb	100
Mo	100
Sn	100
Ti	100

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000 μ g/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-926-3000 in the USA, +31-118-67-1000 in Europe or visit our web-site at www.varianinc.com