



Agilent Technologies

www.agilent.com

Expiry: _____

Certificate of Analysis

Part Number: 6610030600

Lot Number: 11H169

Shelf Life: 18 Months

Calibration Mix 2
5% HNO₃

Concentrations in ug/mL \pm 0.5%

Al	100	Ag	100
As	100	Tl	100
Ba	100	Th	100
Be	100	U	100
Cd	100	V	100
Cr	100	Zn	100
Co	100		
Cu	100		
Pb	100		
Mn	100		
Ni	100		
Se	100		

This standard solution was prepared using high-purity starting materials, high-purity acid (if required) and 18-megaohm de-ionized water. The starting materials were weighed to five significant figures and diluted in volumetric glassware calibrated to five significant figures.

Starting materials were analyzed at 1000 μ g/mL by ICP-MS for trace impurities. The standard solution concentrations were certified instrumentally against the National Institute of Standards and Technology's SRM 3100 series, NIST approved second source and/or gravimetrically.

Accuracy and stability are guaranteed to within plus or minus 0.5% of the certified value for the stated shelf life from the date of shipment. The solution should be kept tightly capped and stored under normal laboratory conditions. See attached MSDS for proper handling information.

For questions or comments please call 1-800-926-3000 in the USA, +31-118-67-1000 in Europe or visit our web-site at www.varianinc.com