

Certificate of Analysis

ISO 17034

Pesticide Standard 4

Product Number: AGI-P03-4**Page:** 1 of 2**Lot Number:** 0006473490**Lot Issue Date:** 28-Jun-2019**Expiration Date:** 31-Jan-2021

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
bromacil	000314-40-9	NT01886	100.4 ± 0.5 µg/mL
cycloate	001134-23-2	RM02853	100.5 ± 0.5 µg/mL
endrin	000072-20-8	RM14459	100.4 ± 0.5 µg/mL
epoxiconazole	133855-98-8	RM13872	100.3 ± 0.5 µg/mL
fenothiocarb	062850-32-2	RM13115	100.4 ± 0.5 µg/mL
hexachlorobenzene	000118-74-1	RM03125	100.4 ± 0.5 µg/mL
iprobefos	026087-47-8	RM13790	100.3 ± 0.5 µg/mL
dicofol	000115-32-2	RM14743	100.4 ± 0.5 µg/mL
quinoxiphen	124495-18-7	RM14287	100.2 ± 0.5 µg/mL
propazine	000139-40-2	RM14562	100.2 ± 0.5 µg/mL
tebupirimphos	096182-53-5	RM12654	100.2 ± 0.5 µg/mL
tetradifon	000116-29-0	RM03451	100.3 ± 0.5 µg/mL
thiobencarb	028249-77-6	RM12027	100.3 ± 0.5 µg/mL
triadimenol	055219-65-3	RM03453	100.3 ± 0.5 µg/mL

Matrix: acetonitrile**Storage:** Store Refrigerated (2° to 8°C)

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

Certificate of Analysis

ISO 17034

Pesticide Standard 4

Product Number: AGI-P03-4

Page: 2 of 2

Lot Number: 0006473490

Lot Issue Date: 28-Jun-2019

Expiration Date: 31-Jan-2021

Analyte

CAS#

Analyte Lot

True Value



Monica Bourgeois
QMS Representative