

# Certificate of Analysis

## ISO 17034

**Product Number:** IAA-240  
**Lot Number:** 0006479624

**Lot Issue Date:** 03-Dec 2018  
**Expiration Date:** 31-Jan 2026

**Product Name:** Zirconium AA Standard

**Description:**

This Reference Material (RM) was gravimetrically prepared in accordance with ISO 17034 and under Agilent's ISO 9001 registered quality system. The neat materials used for this product have been verified by Agilent's ISO 17025 laboratory and under Agilent's ISO 17034 accreditation. The analyte concentrations were verified by Agilent's ISO 17025 accredited laboratory. For each analyte, the true value, with its uncertainty value calculated at the 95% confidence level, is reported below.

Analyte	Starting Material	Lot Number	Purity (%)	Analyte Concentration	Traceability & Method
zirconium	zirconium oxynitrate hydrate	RM10631	99.995	1000 ± 5 µg/mL	NIST SRM 3169; ICP-OES

**Solvent:** 5% nitric acid in low TOC water (< 50 ppb)

**Storage:** Store at Room Temperature (15° to 30°C).

**Traceability:**

Traceability has been established through an unbroken chain of comparisons, each having stated uncertainties. Comparisons are based on appropriate physical or chemical measurements, including gravimetric or volumetric dilution, where the mass or volume of a solution before and after dilution is measured. The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 819.

**Estimation of Uncertainties:**

The true value is reported, with its uncertainty value calculated at the 95% confidence level.

**Homogeneity:**

This RM was formulated and unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Intended Use:**

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods and continuing calibration verification.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening and should be processed without delay for the true value to be valid within the stated uncertainties. Do not pipet from the bottle. Do not return any material removed for pipetting to the bottle. Tightly cap the bottle after removing any material and store according to the instructions noted above.

**Hazards:**

Refer to the Safety Data Sheet for information regarding this RM.

**Expiration of Certification:**

The certification of this RM is valid, within the measurement uncertainty specified, until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.



# Certificate of Analysis ISO 17034

**Product Number:** IAA-240  
**Lot Number:** 0006479624

**Lot Issue Date:** 03-Dec 2018  
**Expiration Date:** 31-Jan 2026

**Maintenance of Certification:**

The real-time, long term stability of the RM may be monitored over the lifetime of the certification. If substantive changes occur that affect the certification before the expiration of this certificate, Agilent Technologies will notify the purchaser.

Monica Bourgeois  
QMS Representative



ISO 17034 Cert No.  
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026  
registered ISO 9001 Quality Management System



ISO 17025 Cert No.  
AT-1937