

Certificate of Analysis

ISO 17034

Pesticides Standard

Product Number: NPM-104A-1**Page:** 1 of 1**Lot Number:** 0006475307**Lot Issue Date:** 27-Jun-2019**Expiration Date:** 31-Jul-2021

This ISO 17034 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
alachlor	015972-60-8	RM14112	1003 ± 5 µg/mL
atraton	001610-17-9	RM00679	1001 ± 5 µg/mL
bromacil	000314-40-9	NT01886	1002 ± 5 µg/mL
butylate	002008-41-5	RM03653	1002 ± 5 µg/mL
chlorpropham	000101-21-3	RM02854	1004 ± 5 µg/mL
hexazinone	051235-04-2	RM13246-004	1003 ± 5 µg/mL
molinate	002212-67-1	RM14835	1004 ± 5 µg/mL
pronamide	023950-58-5	RM14691	1003 ± 5 µg/mL
tetrachlorvinphos	022248-79-9	RM05693	1004 ± 5 µg/mL
tricyclazole	041814-78-2	RM03380	1003 ± 5 µg/mL

Matrix: acetone**Storage:** Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS RepresentativeISO 17034 Cert No.
AR-1936Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management SystemISO17025 Cert No.
AT-1937