

Certificate of Analysis

ISO Guide 34

Pesticides Standard

Product Number: SCA-200-1

Page: 1 of 2

Lot Number: 0006460709

Lot Issue Date: 30-Apr-2019

Expiration Date: 31-May-2020

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
abamectin	071751-41-2	RM13595	100.1 ± 0.5 µg/mL
acephate	030560-19-1	RM14041	100.1 ± 0.5 µg/mL
acequinocyl	057960-19-7	RM11787	100.1 ± 0.5 µg/mL
acetamiprid	135410-20-7	RM11002	100.4 ± 0.5 µg/mL
aldicarb	000116-06-3	r0000173-1	100.1 ± 0.5 µg/mL
azoxystrobin	131860-33-8	RM03833	100.1 ± 0.5 µg/mL
bifenazate	149877-41-8	RM13739	100.4 ± 0.5 µg/mL
bifenthrin	082657-04-3	RM14278	100.2 ± 0.5 µg/mL
boscalid	188425-85-6	RM09348	100.4 ± 0.5 µg/mL
carbaryl	000063-25-2	RM09161	100.3 ± 0.5 µg/mL
carbofuran	001563-66-2	RM12282	100.4 ± 0.5 µg/mL
chlorfenapyr	122453-73-0	RM11763	100.1 ± 0.5 µg/mL
chlorpyrifos	002921-88-2	RM14246	100.3 ± 0.5 µg/mL
clofentezine	074115-24-5	RM11764	100.1 ± 0.5 µg/mL
cyfluthrin	068359-37-5	RM09294	100.2 ± 0.5 µg/mL
cypermethrin	052315-07-8	RM11368	100.1 ± 0.5 µg/mL
diazinon	000333-41-5	RM09202	100.2 ± 0.5 µg/mL
daminozide	001596-84-5	RM09968	100.3 ± 0.5 µg/mL
chlorantraniliprole	500008-45-7	RM14473	100.3 ± 0.5 µg/mL
dichlorvos	000062-73-7	RM11064	100.4 ± 0.5 µg/mL

Matrix: acetonitrile

Certificate of Analysis

ISO Guide 34

Pesticides Standard

Product Number: SCA-200-1

Page: 2 of 2

Lot Number: 0006460709

Lot Issue Date: 30-Apr-2019

Expiration Date: 31-May-2020

Analyte


CAS#

Analyte Lot

True Value

Storage: Store Frozen (-25° to -10°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



Monica Bourgeois

QMS Representative



ISO Guide 34 Cert No.
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System



ISO17025 Cert No.
AT-1937