

Certificate of Analysis

ISO Guide 34

Supplemental Cannabis Pesticide Standard

Product Number: SCA-211-1

Page: 1 of 1

Lot Number: CT-1517

Lot Issue Date: 25-Mar-2019

Expiration Date: 30-Apr-2020


This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent Technologies ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
captan	000133-06-2	RM02848	100.2 ± 0.5 µg/mL
coumaphos	000056-72-4	RM14165	100.4 ± 0.5 µg/mL
dimethomorph	110488-70-5	RM14353	100.4 ± 0.5 µg/mL
fenhexamid	126833-17-8	RM11815	100.3 ± 0.5 µg/mL
pentachloronitrobenzene	000082-68-8	RM12207	100.1 ± 0.5 µg/mL
mevinphos	007786-34-7	RM14344	100.2 ± 0.5 µg/mL
spinetoram	187166-40-1	RM14011	100.1 ± 0.5 µg/mL

Matrix: acetonitrile

Storage: Store Frozen (-25° to -10°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS Representative



ISO Guide 34 Cert No.
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System



ISO17025 Cert No.
AT-1937