

Certificate of Analysis

ISO Guide 34

Pesticides Calibration Standard

Product Number: US-360-1

Page: 1 of 1

Lot Number: 0006462361

Lot Issue Date: 03-May-2019

Expiration Date: 31-May-2020


This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
norflurazon	027314-13-2	RM03744	100.1 ± 0.5 µg/mL
alachlor	015972-60-8	RM14112	100.1 ± 0.5 µg/mL
diphenamid	000957-51-7	RM14708	100.1 ± 0.5 µg/mL
pebulate	001114-71-2	RM12079	100.1 ± 0.5 µg/mL
fenarimol	060168-88-9	RM13668	100.2 ± 0.5 µg/mL
vernolate	001929-77-7	RM03471	100.3 ± 0.5 µg/mL
metribuzin	021087-64-9	RM14478	100.2 ± 0.5 µg/mL
triadimefon	043121-43-3	RM12623	100.3 ± 0.5 µg/mL
butylate	002008-41-5	RM03653	100.1 ± 0.5 µg/mL
simazine	000122-34-9	RM00622	100.3 ± 0.5 µg/mL
simetryn	001014-70-6	RM11761	100.3 ± 0.5 µg/mL
prometryn	007287-19-6	RM14652	100.3 ± 0.5 µg/mL

Matrix: acetone

Storage: Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.


Monica Bourgeois
QMS Representative



ISO Guide 34 Cert No.
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026
registered ISO 9001 Quality Management System



ISO17025 Cert No.
AT-1937