



Certificate of Analysis

Headspace OQ/PV Standard Kit

**Agilent Part
Number:** 5182-9733

**Sample Lot
Number:** N082208

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Item	Lot #	Analyte	CAS#	Certified Value
5182-9733-1	CP-3716	nitrobenzene	000098-95-3	2.01 g/L ($\pm 0.5\%$)
		1,2-dichlorobenzene	000095-50-1	2.01 g/L ($\pm 0.5\%$)
		tert-butyl disulfide	000110-06-5	2.01 g/L ($\pm 0.5\%$)

Solvent: ethanol

**MICROCAPS
Solvent:** N/A

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


Purities:

nitrobenzene	99%
1,2-dichlorobenzene	99%
tert-butyl disulfide	99%
ethanol (absolute)	99%

Date of release: 03 August 2016

Date of expiration: 31 August 2018


John Russo
President


Monica Bourgeois
Director of QA/RA