



Certificate of Analysis

Veterinary Drugs Comprehensive Mix – Submix 7

**Agilent Part
Number:** 5190-0575

**Sample Lot
Number:** CL-0520

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Certified Value
Dimetridazole, hydroxy (HMMNI)	000936-05-0	RM06089	100.4 ± 0.5 µg/mL
Iprnidazole	014885-29-1	RM05942	100.0 ± 0.5 µg/mL
Iprnidazole, hydroxy (IPZOH)	035175-14-5	RM06078	100.3 ± 0.5 µg/mL
Metronidazole	000443-48-1	RM03202	100.0 ± 0.5 µg/mL
Metronidazole, hydroxy (MNZOH)	004812-40-2	RM06022	99.9 ± 0.5 µg/mL
Dimetridazole	000551-92-8	RM04618	100.0 ± 0.5 µg/mL
Ronidazole	007681-76-7	RM06008	100.1 ± 0.5 µg/mL

Solvent: acetonitrile

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Storage: Store at <4°C, Do Not Freeze

Date of release: 12 February 2014

Date of expiration: 31 March 2015


William J. Lea
Quality Assurance Manager