



Agilent Technologies



Certificate of Analysis

Agilent HPLC Flushing Solvent

**Agilent Part
Number:** G1969-85026

**Sample Lot
Number:** CH-0084

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	True Value
isopropyl alcohol	000067-63-0	DC624	50 % (v/v)
acetonitrile	000075-05-8	DC582	25 % (v/v)
cyclohexane	000110-82-7	DB938	15 % (v/v)
dichloromethane	000075-09-2	DB678	10 % (v/v)

Solvent: none

ULTRA Scientific uses calibrated Class A glassware and clean bottles in the manufacture of ULTRA standards. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: 13 January 2011
Date of expiration: 28 February 2015


William J. Leach
Quality Assurance Manager