

# Declaration of Conformity to Manufacturing Specifications

**Product:** D1000 ScreenTape


**Part No:** 5067-5582

The product above has been manufactured using a ISO 9001:2015 certified quality assurance system and tested in accordance to ISO 2859-1.

## Kit Information

During verification, the following characteristics were tested for conformance to internal specifications.

Item	Method	Result
ScreenTape integrity	Visual inspection	<input checked="" type="checkbox"/> passed
Fragment migration behavior*	Electrophoretical	<input checked="" type="checkbox"/> passed
Sizing accuracy in accordance with specification*	Electrophoretical	<input checked="" type="checkbox"/> passed
Sizing precision in accordance with specification*	Electrophoretical	<input checked="" type="checkbox"/> passed
Resolution in accordance with specification*	Electrophoretical	<input checked="" type="checkbox"/> passed
*Sampling in accordance with ISO2859-1, tested with DNA Reference Ladder.		

<b>Lot No:</b>	0201846
<b>Expiration Date:</b>	2021-06-22
<b>Release Date:</b>	2020-09-19
Signed:  Joann Schuelke, QA Manager	
Agilent Technologies Manufacturing GmbH & Co. KG Hewlett-Packard-Str. 8 76337 Waldbronn Germany	

## Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.