

## Declaration of Conformity to Manufacturing Specifications

**Product:** RNA ScreenTape


**Part No:** 5067-5576

The product above has been manufactured and tested in accordance with an ISO 9001:2015 certified quality assurance system.

### Kit Information

During verification, the following characteristics were tested for conformance to internal specifications.

Item	Method	Result
ScreenTape integrity	Visual inspection	<input checked="" type="checkbox"/> passed
Fragment migration behavior*	Electrophoretical	<input checked="" type="checkbox"/> passed
* tested with Reference Sample.		

<b>Lot No:</b>	0201920
<b>Expiration Date:</b>	2021-05-25
<b>Release Date:</b>	2020-11-21
Signed:  Joann Schuelke, QA Manager	
Agilent Technologies Manufacturing GmbH & Co. KG Hewlett-Packard-Str. 8 76337 Waldbronn Germany	

### Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.