

## Declaration of Conformity to Manufacturing Specifications

**Product:** TapeStation Test Tape

**Part No:** 5067-5601


The product above has been manufactured using an ISO 9001:2015 certified quality assurance system.

### Kit Information

During verification, the following characteristics were tested for conformance to internal specifications.

Item	Method	Result
ScreenTape Integrity	Visual inspection	<input checked="" type="checkbox"/> passed
Barcode Integrity	Read Test	<input checked="" type="checkbox"/> passed
Fiducial Features	Position Check	<input checked="" type="checkbox"/> passed
Whiteness Check	Spectral Analysis	<input checked="" type="checkbox"/> passed
Electric Features	Continuity Test	<input checked="" type="checkbox"/> passed

<b>Tape ID:</b>	01-S034-230315-01-000093
<b>Expiration Date:</b>	2026-03-14
<b>Release Date:</b>	2023-03-15

 Signed: Joann Schuelke, QA Manager
Agilent Technologies Deutschland GmbH Hewlett-Packard-Str. 8 76337 Waldbronn Germany

### Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.