



5067-5584



## Declaration of Conformity to Manufacturing Specification

<b>Agilent part number:</b>	5067-5584	<b>Lot number:</b>	0203559
<b>Product name:</b>	High Sensitivity D1000 ScreenTape	<b>Expiration date:</b>	2025-09-30
<b>Storage conditions:</b>	2 °C to 8 °C	<b>Release date:</b>	2024-12-28
<b>Shipping conditions:</b>	2 °C to 8 °C	<b>Approved by:</b>	<i>Sebastian Pallmann</i>
		<b>Date:</b>	2024-12-28

Sebastian Pallmann  
 Senior Quality Engineer  
 Agilent Technologies  
 Deutschland GmbH

This product was manufactured and verified in accordance with an ISO 9001 certified quality management system and tested in accordance to ISO 2859-1 as applicable. The manufacture, inspection, and testing were performed by trained personnel. The equipment and materials used in the manufacturing process are identifiable and traceable. All associated test equipment is maintained and calibrated.

The following product characteristics were tested for conformity to manufacturing specifications.

Product characteristic	Method	Result
ScreenTape integrity	Visual inspection	Pass
Fragment migration behavior	Electrophoretical with DNA reference ladder	Pass
Sizing accuracy	Electrophoretical with DNA reference ladder	Pass
Sizing precision	Electrophoretical with DNA reference ladder	Pass
Resolution	Electrophoretical with DNA reference ladder	Pass

**For Research Use Only. Not for use in diagnostic procedures.**

Agilent Technologies  
 Deutschland GmbH  
 Hewlett-Packard-Str. 8  
 76337 Waldbronn, Germany

**Made in Germany**

**For global support: [www.agilent.com/en/contact-us/page](http://www.agilent.com/en/contact-us/page)**

Template Number: 5067-5584 - DOC MS  
 Template Revision: B.00