



5067-5579



Declaration of Conformity to Manufacturing Specification

| | | | |
|-----------------------------|---------------------------------|-------------------------|---------------------------|
| Agilent part number: | 5067-5579 | Lot number: | 0203666 |
| Product name: | High Sensitivity RNA ScreenTape | Expiration date: | 2026-04-07 |
| Storage conditions: | 2 °C to 8 °C | Release date: | 2025-04-05 |
| Shipping conditions: | 2 °C to 8 °C | Approved by: | <i>Sebastian Pallmann</i> |

Date: 2025-04-05

Sebastian Pallmann
 Senior Quality Engineer
 Agilent Technologies
 Deutschland GmbH

This product was manufactured and verified in accordance with an ISO 9001 certified quality management system and tested in accordance to ISO 2859-1 as applicable. The manufacture, inspection, and testing were performed by trained personnel. The equipment and materials used in the manufacturing process are identifiable and traceable. All associated test equipment is maintained and calibrated.

The following product characteristics were tested for conformity to manufacturing specifications.

| Product characteristic | Method | Result |
|-----------------------------|---|--------|
| ScreenTape integrity | Visual inspection | Pass |
| Fragment migration behavior | Electrophoretical with DNA reference ladder | Pass |

For Research Use Only. Not for use in diagnostic procedures.

Agilent Technologies
 Deutschland GmbH
 Hewlett-Packard-Str. 8
 76337 Waldbronn, Germany

Made in Germany

For global support: www.agilent.com/en/contact-us/page

Template Number: 5067-5579 - DOC MS
 Template Revision: B.00