5067-5582

# Declaration of Conformity to Manufacturing Specification

**Agilent part number:** 5067-5582

**Product name:** D1000 ScreenTape

**Storage conditions:** 2 °C to 8 °C

**Shipping conditions:** 2 °C to 8 °C

**Lot number:** 0203677

**Expiration date:** 2026-01-20

**Release date:** 2025-04-19

**Approved by:**

Sebastian Pallmann
Senior Quality Engineer
Agilent Technologies
Deutschland GmbH

**Date:** 2025-04-19

This product was manufactured and verified in accordance with an ISO 9001 certified quality management system and tested in accordance to ISO 2859-1 as applicable. The manufacture, inspection, and testing were performed by trained personnel. The equipment and materials used in the manufacturing process are identifiable and traceable. All associated test equipment is maintained and calibrated.

The following product characteristics were tested for conformity to manufacturing specifications.

| Product characteristic | Method | Result |
|---|---|---|
| ScreenTape integrity | Visual inspection | Pass |
| Fragment migration behavior | Electrophoretical with DNA reference ladder | Pass |
| Sizing accuracy | Electrophoretical with DNA reference ladder | Pass |
| Sizing precision | Electrophoretical with DNA reference ladder | Pass |
| Resolution | Electrophoretical with DNA reference ladder | Pass |

**For Research Use Only. Not for use in diagnostic procedures.**

Agilent Technologies
Deutschland GmbH
Hewlett-Packard-Str. 8
76337 Waldbronn, Germany

**Made in Germany**

**For global support: www.agilent.com/en/contact-us/page**