



5067-5579



## Declaration of Conformity to Manufacturing Specification

|                             |                                 |                         |                           |
|-----------------------------|---------------------------------|-------------------------|---------------------------|
| <b>Agilent part number:</b> | 5067-5579                       | <b>Lot number:</b>      | 0203988                   |
| <b>Product name:</b>        | High Sensitivity RNA ScreenTape | <b>Expiration date:</b> | 2026-12-19                |
| <b>Storage conditions:</b>  | 2 °C to 8 °C                    | <b>Release date:</b>    | 2025-12-20                |
| <b>Shipping conditions:</b> | 2 °C to 8 °C                    | <b>Approved by:</b>     | <i>Sebastian Pallmann</i> |

**Date:** 2025-12-20

Sebastian Pallmann  
 Quality Manager  
 Agilent Technologies  
 Deutschland GmbH

This product was manufactured and verified in accordance with an ISO 9001 certified quality management system and tested in accordance to ISO 2859-1 as applicable. The manufacture, inspection, and testing were performed by trained personnel. The equipment and materials used in the manufacturing process are identifiable and traceable. All associated test equipment is maintained and calibrated.

The following product characteristics were tested for conformity to manufacturing specifications.

| Product characteristic      | Method                                  | Result |
|-----------------------------|---|--------|
| ScreenTape integrity        | Visual inspection                       | Pass   |
| Fragment migration behavior | Electrophoretical with reference sample | Pass   |

**For Research Use Only. Not for use in diagnostic procedures.**

Agilent Technologies  
 Deutschland GmbH  
 Hewlett-Packard-Str. 8  
 76337 Waldbronn, Germany

**Made in Germany**

**For global support: [www.agilent.com/en/contact-us/page](http://www.agilent.com/en/contact-us/page)**

Template Number: 5067-5579 - DOC MS  
 Template Revision: C.00