



5067-5576



Declaration of Conformity to Manufacturing Specification

Agilent part number:	5067-5576	Lot number:	0204059
Product name:	RNA ScreenTape	Expiration date:	2027-02-13
Storage conditions:	2 °C to 8 °C	Release date:	2026-02-14
Shipping conditions:	2 °C to 8 °C	Approved by:	<i>Sebastian Pallmann</i>

Date: 2026-02-14

Sebastian Pallmann
 Quality Manager
 Agilent Technologies
 Deutschland GmbH

This product was manufactured and verified in accordance with an ISO 9001 certified quality management system and tested in accordance to ISO 2859-1 as applicable. The manufacture, inspection, and testing were performed by trained personnel. The equipment and materials used in the manufacturing process are identifiable and traceable. All associated test equipment is maintained and calibrated.

The following product characteristics were tested for conformity to manufacturing specifications.

Product characteristic	Method	Result
ScreenTape integrity	Visual inspection	Pass
Fragment migration behavior	Electrophoretical with reference sample	Pass

For Research Use Only. Not for use in diagnostic procedures.

Agilent Technologies
 Deutschland GmbH
 Hewlett-Packard-Str. 8
 76337 Waldbronn, Germany

Made in Germany

For global support: www.agilent.com/en/contact-us/page

Template Number: 5067-5576 - DOC MS
 Template Revision: C.00