



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch No. K07-B8160

Part Number 5190-1526

Description Syringe, 500 uL PTFE RN LC tip

Accuracy: $\pm 1\%$

Reproducibility: $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.