



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch No. K09-B9094

Part Number 5190-1534

Description Syringe, 2.5 mL, PTFE, luer lok

Accuracy: $\pm 1\%$

Reproducibility: $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.