



**Agilent Technologies**

**Certificate of Conformance**

Product: Syringe

Batch No. Q06-D3727

Part Number 5190-1534

Description Syringe, 2.5 mL, PTFE, luer lok

Accuracy:  $\pm 1\%$

Reproducibility:  $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.