



# Agilent Technologies

## *Certificate of Conformance*

Product: Syringe

Batch No. M03-C0884

Part Number 5190-1542

Description Syringe, 10 mL PTFE RN bevel tip

Accuracy:  $\pm 1\%$

Reproducibility:  $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.