



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch Code	H11A1584
Part Number	5190-1471
Description	Syringe, 2.0 uL RN 23 g cone tip

Accuracy:	$\pm 2\%$
Reproducibility:	$\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.