

## **Certificate of Conformance**

Product: Syringe

Batch Code H04C9608

Part 5190-1479

Description Syringe, 5.0 uL FN 26 g bevel tip

Accuracy: ±1%

Reproducibility: ±1%

This syringe has been manufactured to the highest quality using measuring equipment and instrumentation calibrated and traceable to International Standards.

Agilent Technologies ensures the displacement is within the stated conformance specifications.