



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch Code	H08D1166
Part Number	5190-1487
Description	Syringe, 10.0 uL, FN, 26 g bevel tip,

Accuracy:	±1%
Reproducibility:	±1%

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.