



**Agilent Technologies**

**Certificate of Conformance**

Product: Syringe

**Batch Code**                      **H12D3095**

**Part Number**                    **5190-1494**

**Description**                    **Syringe, 25.0 uL, FN, LC tip**

Accuracy:                          $\pm 1\%$

Reproducibility:                 $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.