



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch Code **H05A8933**

Part **5190-1520**

Description **Syringe, 250 uL PTFE RN LC tip**

Accuracy: $\pm 1\%$

Reproducibility: $\pm 1\%$

This syringe has been manufactured to the highest quality using measuring equipment and instrumentation calibrated and traceable to International Standards.

Agilent Technologies ensures the displacement is within the stated conformance specifications.