



**Agilent Technologies**

**Certificate of Conformance**

Product: Syringe

<b>Batch Code</b>	<b>H09D1827</b>
<b>Part Number</b>	<b>5190-1521</b>
<b>Description</b>	<b>Syringe, 500 uL, FN, bevel tip</b>

Accuracy:	±1%
Reproducibility:	±1%

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.