



**Agilent Technologies**

**Certificate of Conformance**

Product: Syringe

|                    |                                    |
|--------------------|------------------------------------|
| <b>Batch Code</b>  | <b>H07D0827</b>                    |
| <b>Part Number</b> | <b>5190-1522</b>                   |
| <b>Description</b> | <b>Syringe, 500 uL, FN, LC tip</b> |

Accuracy:  $\pm 1\%$

Reproducibility:  $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.