



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch Code **H08A9682**

Part Number **5190-1525**

Description **Syringe, 500 uL PTFE RN bevel tip**

Accuracy: $\pm 1\%$

Reproducibility: $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.