



**Agilent Technologies**

## **Certificate of Conformance**

*Product: Syringe*

**Batch Code**    *H05A8845*

**Part**            *5190-1533*

**Description**    *Syringe, 2.5 mL PTFE RN bevel tip*

*Accuracy:*             $\pm 1\%$

*Reproducibility:*     $\pm 1\%$

*This syringe has been manufactured to the highest quality using measuring equipment and instrumentation calibrated and traceable to International Standards.*

*Agilent Technologies ensures the displacement is within the stated conformance specifications.*