



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch Code **H06A9237**

Part Number **5190-1543**

Description **Syringe, 10 mL, PTFE,luer lok valv**

Accuracy: $\pm 1\%$

Reproducibility: $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.