



Agilent Technologies

Certificate of Conformance

Product: Syringe

Batch Code **H10B0237**

Part Number **5190-1545**

Description **Syringe, 25 mL, PTFE, luer lok va**

Accuracy: $\pm 1\%$

Reproducibility: $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.