



**Agilent Technologies**

**Certificate of Conformance**

Product: Syringe

<b>Batch Code</b>	<b>H08A9710</b>
<b>Part Number</b>	<b>5190-1546</b>
<b>Description</b>	<b>Syringe, 50 mL PTFE luer lok</b>

Accuracy:  $\pm 1\%$

Reproducibility:  $\pm 1\%$

This syringe has been manufactured to the highest quality using in-process measuring equipment and instrumentation calibrated and traceable to International Standards through the National Research Laboratory of Metrology (NRLM).

Agilent Technologies ensures the displacement is within the stated conformance specifications.