Declaration of Conformity to Manufacturing Specifications

Product: Agilent High Sensitivity DNA Chips
Part No: G2938-68000
Included in Part No: 5067-4626
Lot No: AP23BK50

Note: The chips were prepared using ISO 9001 quality processes. **For Research Use Only. Not for use in diagnostic procedures.**

Kit Information

During specification verification, the following characteristics were tested on a statistically significant number of chips:

Specification	Method	Result
Channel Depth	Profilometer	Passed ☑
Optical Inspections	Visual Inspections	Passed ☑

Kit Expiration Date:	23.FEB.2023

Date: 07.MAR.2022

Signed: S.Knapp

PerkinElmer, Inc.
Hopkinton, MA 01748

Agilent Technologies